



# Unique Device Identifier Timelines

Compliance Date	Requirements
<b>September 24, 2014*</b> Labels and packages of: <b>Class III devices [§801.20], plus...</b>	<ul style="list-style-type: none"><li>• Class III stand alone software [§801.50(b)]</li><li>• Devices licensed under the Public Health Service Act [§801.20]</li><li>• Dates on labels must be formatted as YYYY-MM-DD [§801.18]**</li><li>• Data for these devices must be submitted to GUDID [§830.300]</li></ul>
<b>September 24, 2015</b> Labels and packages of: <b>Implantable, life-supporting &amp; life-sustaining devices [§801.20], plus...</b>	<ul style="list-style-type: none"><li>• Life-supporting/life-sustaining Stand-Alone Software must have UDI [§801.50(b)]</li><li>• Life-supporting/life-sustaining devices must have UDI as permanent mark if it is to be used more than once and reprocessed before each use [§801.45]</li><li>• Dates on labels must be formatted as YYYY-MM-DD [§801.18]**</li><li>• Data for these devices must be submitted to GUDID [§830.300]</li></ul>
<b>September 24, 2016</b> Labels and packages of: <b>Class II devices &amp; Software [§801.20], plus...</b>	<ul style="list-style-type: none"><li>• Class III devices must have UDI as permanent mark if it is to be used more than once and reprocessed before each use [§801.45]</li><li>• Dates on labels must be formatted as YYYY-MM-DD [§801.18]**</li><li>• Data for these devices must be submitted to GUDID [§830.300]</li></ul>
<b>September 24, 2018</b> Labels and packages of: <b>Class I devices &amp; Software [§801.20], plus...</b>	<ul style="list-style-type: none"><li>• Class II devices must have UDI as permanent mark if it is to be used more than once and reprocessed before each use [§801.45]</li><li>• Devices that have not been classified as Class I, Class II or Class III [§801.20]</li><li>• Dates on labels of ALL devices, including devices exempted from UDI labeling requirements must be in YYYY-MM-DD format [§801.18]**</li><li>• Data for these devices must be submitted to GUDID</li></ul>
<b>September 24, 2020</b> <b>The final hurrah...</b>	<ul style="list-style-type: none"><li>• Class I devices and devices that have not been classified as Class I, Class II or Class III must have UDI as permanent mark if it is to be used more than once and reprocessed before each use [§801.45]</li></ul>

\*June 23, 2014 Request 1-year extension